## DP-50 HIGH VOLTAGE DIFFERENTIAL PROBE-

## **SPECIFICATIONS**

(1) Bandwidth: DC - to 50 MHz (-3 dB)

(2) Attenuation: X100, X200, X500, X1000

(3) Accuracy:  $\pm 2\%$ 

(4) Voltage Input Ranges (DC + AC peak to peak)

 $\leq$  700 Vp-p for X100, (i.e about 230 V RMS or DC)

 $\leq$  1400 Vp-p for X200, (i.e about 460 V RMS or DC)

 $\leq$  3500 Vp-p for X500, (i.e about 1140 V RMS or DC)

 $\leq$  7000 Vp-p for X1000, (i.e about 2300 V RMS or DC)

(5) Permitted Max Input Voltage

Max differential voltage: 7000 V (DC + AC peak to peak)
Max voltage between each input terminal and ground:
1200 V RMS

(6) Input Impedance:

Differential:  $16 M\Omega / 1.2 pF$ 

Between terminals and ground:  $8 M\Omega / 2.3 pF$ 

(7) Output:  $\leq \pm 7.0 \text{ V}$ 

(8) Output Impedance: 50  $\Omega$ 

(9) Rise Time: 7 ns

(10) Rejection Rate on Common Mode:

 $60 \text{ Hz:} > 80 \text{ dB} \; ; \; 100 \text{ Hz:} > 60 \text{ dB} \; ; \; 1 \text{ MHz:} > 50 \text{ dB}$ 

(11) Power Supply: Only External 9 V DC power supply.

(12) Consumption: < 300 mA max